

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

CB TEST CERTIFICATE

Product

Ultrasonic Flow and Energy Meter

Name and address of the applicant

Badger Meter
8635 Washington Avenue
Racine, WI 53406
United States of America

Name and address of the manufacturer

Badger Meter
8635 Washington Avenue
Racine, WI 53406
United States of America

Name and address of the factory

Note: When more than one factory, please report on page 2

☐ Additional information on page 2Badger Meter
8635 Washington Avenue
Racine, WI 53406
United States of America

Ratings and principal characteristics

10-28VDC, 40W, powered from external 100-240VAC power supply

Trademark (if any)

RFI

Customer's Testing Facility (CTF) Stage used

Model / Type Ref.

Dynasonics model DXNP-abS-NN, where a = power cord option and b= sensor kit option

Additional information (if necessary may also be reported on page 2)

☐ Additional information on page 2

A sample of the product was tested and found to be in conformity with

IEC 61010-1:2001

National differences:

EU Group Differences, AT, AU, BE, CA, CH, CZ, DE, DK, FI, FR, GB, HU, IL, IT, KR, NL, NO, RU, SE, SG, SK, SL, UA, US

As shown in the Test Report Ref. No. which forms part of this Certificate

US 945 MET M0A0

This CB Test Certificate is issued by the National Certification Body

MET Laboratories, Inc.
914 W. Patapsco Avenue, Baltimore MD 21230-3432, United States of America